

TECHNICAL BULLETIN #1

May 31, 2011

To: All EMxpert and RFxpert end-users
Re: Agilent IO 16.1

Agilent has discontinued its support for IO Library Suite 16.0 and lower.

RFxpert and EMxpert installation packages will ship effective May 31, 2011 with Agilent 16.1.

Known limitations are as follows:

- Windows XP 32-Bit requires SP3
- Windows XP 64-Bit is not supported
- Windows Vista 32-Bit requires SP1
- Windows Vista 64-Bit requires SP1
- Windows 7 32-Bit works fine as is. No SP currently available.
- Windows 7 64-Bit requires all the latest OS updates. It will not work without the latest Windows updates.

You can download IO 16.1 and upgrade immediately your current EMSCAN applications:
http://aaa.product.agilent.com/IOLibraries/Download_IO_Libraries.aspx